



Pan Melanoma + S100 (Tyrosinase + MART-1 + S100)

ISO
9001:2000
CERTIFIED

Prediluted Double Stain Antibody (4-Step)
Control Number: 901-213DS-032309

Catalog Number: PPM 213 DS AA, H
Description: 6.0, 25 ml, prediluted
Dilution: Ready-to-use
Diluent: N/A

Intended Use:
For In Vitro Diagnostic Use

Summary and Explanation:

Tyrosinase is a key enzyme involved in the initial stages of melanin biosynthesis. Studies have shown Tyrosinase to be a more sensitive marker when compared to HMB45 and MART-1. It has also shown to label a higher percentage of desmoplastic melanomas than HMB45.

MART-1 Cocktail (M2-7C10 + M2-9E3) recognizes a protein of 18kDa, identified at MART-1 (Melanoma Antigen Recognized by T cells 1). MART-1 is a useful addition to melanoma panels as it is apparently specific for melanocytic lesions. Studies have also shown that MART-1 is more sensitive than HMB45 when labeling metastatic melanomas. This MART-1 cocktail does not stain steroid tumors unlike Melan A [103].

S100 stains Schwannomas, ependymomas, astroglomas, almost all benign and malignant melanomas and their metastases. S100 protein is also expressed in the antigen presenting cells such as the Langerhans cells in skin and interdigitating reticulum cells in the paracortex of lymph nodes.

Principle of Procedure:

Antigen detection, in tissues and cells, is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a universal, affinity-purified, secondary antibody is added to bind to the primary antibody. An enzyme label is then added to bind to the secondary antibody; this detection of the bound antibody is evidenced by a colorimetric reaction.

Source: Mouse monoclonal and Rabbit polyclonal

Species Reactivity: Human; others not tested.

Clone: T311 + M2-7C10 + M2-9E3+N/A

Isotype: IgG_{2a} + IgG_{2b/kappa} + IgG_{2b/kappa} + N/A

Epitope/Antigen: Tyrosinase, MART-1, S100 protein

Cellular Localization:

Tyrosinase: cytoplasmic (brown), Mart-1: cytoplasmic (brown) S100 nuclear/cytoplasmic (red)

Positive Control: Melanoma

Normal Tissue: Skin, nevus

Abnormal Tissue: Melanoma

Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

Supplied As: Buffer with protein carrier and preservative.

Storage and Stability:

Store at 2°C to 8°C. Do not use after expiration date printed on vial. If reagents are stored under conditions other than those specified in the package insert, they must be verified by the user. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

Peroxide Block:

Block for 5 minutes with BIOCARE's PEROXIDAZED 1.

Pretreatment Solution (recommended): Diva or Reveal

Pretreatment Protocol:

Heat Retrieval Method:

Retrieve sections under pressure using BIOCARE's Decloaking Chamber, followed by a wash in distilled water. Alternatively, steam tissue sections for 45-60 minutes. Allow solution to cool for 20 minutes then wash in distilled water.

Protein Block:

Incubate for 10-15 minutes at RT with BIOCARE's Background Sniper.

Primary Antibody: Incubate for 30-60 minutes at RT.

Double Stain Detection:

Incubate for 30 minutes at RT using BIOCARE's Double Stain Kit #2.

Chromogen (1): Incubate for 5 minutes at RT when using BIOCARE's Betazoid DAB.

Chromogen (2):

Incubate for 10-15 minutes at RT with BIOCARE's Vulcan Fast Red.

Counterstain:

Rinse with deionized water. Incubate for 30-60 seconds with Tacha's Automated Hematoxylin. Rinse with deionized water. Apply Tacha's Bluing solution for 1 minute.

Technical Note:

This antibody has been standardized with BIOCARE's Double Stain Kit #2. It can also be used on an automated staining system. Use TBS buffer for washing steps.

Performance Characteristics:

The optimum antibody dilution and protocols for a specific application can vary. These include, but are not limited to: fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of BIOCARE products. Ultimately, it is the responsibility of the investigator to determine optimal conditions. These products are tools that can be used for interpretation of morphological findings in conjunction with other diagnostic tests and pertinent clinical data by a qualified pathologist.

Quality Control:

Refer to NCCLS Quality Assurance for Immunocytochemistry approved guidelines, December 1999 MM4-A Vol.19 No.26 for more information about Tissue Controls.

Precautions:

This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC.

Sodium azide (NaN₃) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for disease control, 1976, National Institute of Occupational Safety and Health, 1976)

Specimens, before and after fixation and all materials exposed to them, should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water.

Microbial contamination of reagents may result in an increase in nonspecific staining. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change. The MSDS is available upon request.

Troubleshooting:

Follow the antibody specific protocol recommendations according to data sheet provided. If atypical results occur, contact BIOCARE's Technical Support at 1-800-542-2002.



**Pan Melanoma + S100 (Tyrosinase + MART-1 + S100)**

Prediluted Double Stain Antibody (4-Step)

Control Number: 901-213DS-032309

ISO
9001:2000
CERTIFIED**Limitations and Warranty:**

There are no warranties, expressed or implied, which extend beyond this description. BIOCARE is not liable for property damage, personal injury, or economic loss caused by this product.

References:

1. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
2. National Committee for Clinical Laboratory Standards(NCCLS). Protection of laboratory workers from infectious diseases transmitted by blood and tissue; proposed guideline. Villanova, PA 1991;7(9). Order code M29-P.



Pan Melanoma + S100 (Tyrosinase + MART-1 + S100)

Prediluted Mouse Monoclonal and Rabbit Polyclonal Double Stain Antibody (4-Step)

Control Number: 901-213DSIP-111909

Catalog Number:

IP 213DS G10
IP 213DS G20

Description:

10 ml, predilute
20 ml, predilute

Intended Use:


For In Vitro Diagnostic Use

Summary and Explanation:

Tyrosinase is a key enzyme involved in the initial stages of melanin biosynthesis. Studies have shown Tyrosinase to be a more sensitive marker when compared to HMB45 and MART-1. It has also shown to label a higher percentage of desmoplastic melanomas than HMB45.

MART-1 Cocktail (M2-7C10 + M2-9E3) recognizes a protein of 18kDa, identified as MART-1 (Melanoma Antigen Recognized by T cells 1). MART-1 is a useful addition to melanoma panels as it is apparently specific for melanocytic lesions. Studies have also shown that MART-1 is more sensitive than HMB45 when labeling metastatic melanomas. This MART-1 cocktail does not stain steroid tumors unlike Melan A [103].

S100 stains Schwannomas, ependymomas, astroglomas, almost all benign and malignant melanomas and their metastases. S100 protein is also expressed in the antigen presenting cells such as the Langerhans cells in skin and interdigitating reticulum cells in the paracortex of lymph nodes.



Melanoma stained with Pan Melanoma + S100 antibodies.

Source: Mouse monoclonal and Rabbit polyclonal

Species Reactivity: Human; others not tested

Clone: T311+M2-7C10+M2-9E3

Isotype: IgG_{2a}+ IgG_{2b/kaappa} + IgG_{2b}

Antibody Category: Melanoma

Epitope/Antigen: Tyrosinase, MART-1, S100 protein

Cellular Localization:

Tyrosinase: cytoplasmic (brown), MART-1: cytoplasmic (brown), S100: nuclear/cytoplasmic (red)

Positive Control: Melanoma

Normal Tissue: Skin, nevus

Abnormal Tissue: Melanoma

Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

Supplied As: Buffer with protein carrier and preservative.

Storage and Stability:

Store at 2°C to 8°C. Do not use after expiration date printed on vial. If reagents are stored under conditions other than those specified in the package insert, they must be verified by the user. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

Protocol Recommendations

Pretreatment Solution (recommended): Diva

Pretreatment Protocol:

Heat Retrieval Method:

Retrieve sections under pressure using BIOCARE's Decloaking Chamber, followed by a wash in distilled water. Alternatively, steam tissue sections for 45-60 minutes. Allow solution to cool for 20 minutes then wash in distilled water.

Peroxide Block:

Block for 5 minutes at RT.

Primary Antibody:

Incubate for 30 minutes at RT.

Double Stain Detection:

Incubate for 15 minutes at RT using Multiplex Secondary Reagent 2.

Chromogen (1):

Incubate for 5 minutes at RT with intelliPATH FLX DAB.

Chromogen (2):

Incubate for 10-20 minutes at RT with intelliPATH FLX Fast Red. Rinse in deionized water.

Counterstain:

Rinse with deionized water. Incubate for 5 min with Automated Hematoxylin. Rinse with TBS buffer for 1 min, followed by a rinse with deionized water.

Quality Statement:

BIOCARE protocols have been standardized using in-house antibodies, detection and accessory reagents for use on the intelliPATH FLX automated stainer. Recommended staining protocols are specified in the datasheet of the antibody of interest. Pre-optimized intelliPATH FLX protocols with preset parameters can be displayed, printed and edited according to the procedure in the operator's manual. Refer to the operator's manual for additional instruction to navigate intelliPATH FLX software and stainer. Use TBS for washing steps unless otherwise specified.

Performance Characteristics:

The optimum antibody dilution and protocols for a specific application can vary. These include, but are not limited to: fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of BIOCARE products. Ultimately, it is the responsibility of the investigator to determine optimal conditions. These products are tools that can be used for interpretation of morphological findings in conjunction with other diagnostic tests and pertinent clinical data by a qualified pathologist.

Quality Control:


Refer to NCCLS Quality Assurance for Immunocytochemistry approved guidelines, December 1999 MM4-A Vol.19 No.26 for more information about Tissue Controls.

Precautions:

This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC.

Sodium azide (NaN₃) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for disease control, 1976, National Institute of Occupational Safety and Health, 1976)

Specimens, before and after fixation and all materials exposed to them, should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water.



Pan Melanoma + S100 (Tyrosinase + MART-1 + S100)

Prediluted Mouse Monoclonal and Rabbit Polyclonal Double Stain Antibody (4-Step)

Control Number: 901-213DSIP-111909

Precautions cont'd:

Microbial contamination of reagents may result in an increase in nonspecific staining. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change. The MSDS is available upon request.

Troubleshooting:

Follow the antibody specific protocol recommendations according to data sheet provided. If atypical results occur, contact BIOCARE's Technical Support at 1-800-542-2002.

Limitations and Warranty:

There are no warranties, expressed or implied, which extend beyond this description. BIOCARE is not liable for property damage, personal injury, or economic loss caused by this product.

References:

1. Orchard G. Evaluation of melanocytic neoplasms: application of a pan-melanoma antibody cocktail. Br J Biomed Sci. 2002; 59(4): 196-202.
2. de Vries TJ, Smeets M, de Graaf R, Hou-Jensen K, Brocker EB, Renard N, Eggermont AM, van Muijen GN, Ruiter DJ. Expression of gp100, MART-1, tyrosinase, and S100 in paraffin-embedded primary melanomas metastases: implications for diagnosis and immunotherapy. A study conducted by the EORTC Melanoma Cooperative Group. J Pathol 2001; 193: 13-20.
3. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
4. National Committee for Clinical Laboratory Standards(NCCLS). Protection of laboratory workers from infectious diseases transmitted by blood and tissue; proposed guideline. Villanova, PA 1991;7(9). Order code M29-P.

